

VERIFICATION OF CALIBRATION REPORT

APR 30 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130113

Supervisor/Operator Performing the Verification Procedure:

Name: Kenneth G. Cox ID #: 3527 Date: April 26, 2010 Bmb
February 28, 2010 5/3/10

A

Agency: Skagway Police Department Phone #: 907 983-2232

Instrument Location Skagway Police Department, 1st & State Street Skagway, Alaska 99840

Alco S/N X173019 X173019 Target Value .082 High Pressure 1000
5/11/10

Alco Test Value Average .079 .079
1st Alco 2nd Alco

Signature [Signature]

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of June, 2010.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

APRIL 26, 2010

OPERATOR'S NAME:

COX/KENNETH/G

OPERATOR'S NUMBER: 3527

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: SKG1

CASE/REPORT: 2010-0877

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: ~~1173019~~ ¹⁷³⁰¹⁹ _{008 8/1/10}

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.60 in

ALCO TARGET	.001	18:49
BLANK TEST	.000	18:50
INTERNAL STANDARD	VERIFIED	18:50
ALCO TV 29.60 in	.079	18:51
BLANK TEST	.000	18:52
SUBJECT SAMPLE	.000	18:52
BLANK TEST	.000	18:53
ALCO TV 29.60 in	.079	18:54
BLANK TEST	.000	18:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

APRIL 26, 2010

TIME 18:39

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD =	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~■